

Examples of results from calibration of standard masses class F<sub>2</sub> 100g

Marking	Mass of standard mass	Extended uncertainty	Maximal permissible error (according to R111 OIML)
F 100	100 g + 0,1 mg	0,1 mg	± 1,5 mg